

KNOP



Declaration of Conformity

Manufacturer: Knop Elektronik
Fabriksvej 20
7600 Struer
Denmark

Declares that the products:

EPI2000, Epilepsy and convulsion alarm with 12 bit transmission
EPI2000P, Epilepsy and convulsion alarm with PowerCode transmission
EPI2000S, Epilepsy and convulsion sensor
RX2004, Receiver
TX2004, One-channel transmitter
PU2002, Blow-activated Switch

complies with the Medical Device Directive 93/42/EEC.

The product meets the requirements in:

EN60601-1-2:1993	Electromagnetic compatibility.
EN60601-1:2004	General requirements for basic safety and essential performance.
DS/EN ISO 14971	Medical equipment Application of risk management to medical devices.

Struer, 6th February, 2008

Jan Knop
Managing Director, Knop Elektronik A/S