

KNOP



Declaration of Conformity

Manufacturer: Knop Elektronik
Fabriksvej 20
7600 Struer
Denmark

Declares that the products:

BW900	Bedwetting alarm
EPI900	Epilepsy and convulsion alarm
EPI900S	Epilepsy and convulsion sensor
KK901/KK901P	Ping Pong contact with transmitter
LAK900	Sound-activated alarm
MD900	Man Down alarm
PC900	Pull-core switch
PIR900	Infrared detector and door detector
PU901/PU901P	Blow contact with transmitter
RP900/901/902	Repeater
RX900-2	2 channel receiver
RX900-2RE	2 channel receiver with relay output
RX900-LED	Receiver with LED
RX900A	Receiver with wristband and LCD
RX900/901B	Receiver with LCD
TX900-1	One-channel transmitter
TX900-A	One-channel wristband transmitter
TX901A	One-channel wristband transmitter
CT901	Controller with LED light
LS901	Light Slave
USB900	Programming tool

Complies with:
93/42/EEC Medical Device Directive
2011/65/EU RoHS Directive

The product meets the requirements in:

DS/EN ISO 14971	Medical equipment Application of risk management to medical devices.
EN60601-1-2:2007	Medical electrical equipment Electromagnetic compatibility.
EN60601-1:2006	Medical electrical equipment General requirements for basic safety and essential performance.
DS/EN 50581:2012	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances

Struer, 1th September, 2018

Jan Knop
Managing Director, Knop Elektronik A/S

KNOP
FABRIKSVEJ 20, 7600 STRUER
TEL: +45 9784 0444

